UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/728,267	12/04/2003	William M. Appling	P-37/500622.20040	7104
26418 7 REED SMITH, I	7590 01/11/200 LLP	EXAMINER		
ATTN: PATEN	T RECORDS DEPAR	KOHARSKI, CHRISTOPHER		
599 LEXINGTON AVENUE, 29TH FLOOR NEW YORK, NY 10022-7650			ART UNIT	PAPER NUMBER
			3763	
SHORTENED STATUTORY PERIOD OF RESPONSE		MAIL DATE	DELIVERY MODE	
3 MON	ITHS	01/11/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

_	_ P
V	7

	Application No.	Applicant(s)				
	10/728,267	APPLING, WILLIAM M.				
Office Action Summary	Examiner	Art Unit				
	Christopher D. Koharski	3763				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status		·				
1) Responsive to communication(s) filed on 23 Ma	arch 2006					
_	action is non-final.					
,						
, — · · ·	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.					
	pario Quayro, 1000 0.2111, 10					
Disposition of Claims						
4) Claim(s) <u>1-25</u> is/are pending in the application.						
4a) Of the above claim(s) is/are withdraw	n from consideration.					
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>1-25</u> is/are rejected.	•					
7) ☐ Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/or	election requirement.					
Application Papers						
9) The specification is objected to by the Examiner.						
10) The drawing(s) filed on is/are: a) acce	epted or b) objected to by the f	Examiner.				
Applicant may not request that any objection to the o	drawing(s) be held in abeyance. See	e 37 CFR 1.85(a).				
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
0.1.11.00.0440						
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign	priority under 35 U.S.C. § 119(a))-(d) or (f).				
a) ☐ All b) ☐ Some * c) ☐ None of:						
 Certified copies of the priority documents 	s have been received.	•				
2. Certified copies of the priority documents	have been received in Applicati	on No				
3. Copies of the certified copies of the prior	ity documents have been receive	ed in this National Stage				
application from the International Bureau	(PCT Rule 17.2(a)).					
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)		(270 140)				
) Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948) Interview Summary (PTO-413) Paper No(s)/Mail Date						
	Information Disclosure Statement(s) (PTO/SB/08) 5) Notice of Informal Patent Application					
Paper No(s)/Mail Date <u>10/12/04,11/01/04</u> .	6) Other:					

DETAILED ACTION

Information Disclosure Statement

The information disclosure statements (IDS) that were submitted on 10/12/2004 and 11/01/2004 are in compliance with the provisions of 37 CFR 1.97. Accordingly, the examiner is considering the information disclosure statements.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

Claims 1-22 are rejected under 35 U.S.C. 102(b) as being anticipate by Chee et al. (5,542,937). Chee et al. discloses a multilumen extruded catheter.

Regarding claims 1-22, Chee et al. discloses a catheter comprising three distinct areas comprising a proximal tube (168) with a first polymer material of a first durometer and amount of radiopaque filler, a distal tube segment (166) with a second material and a different amount of radiopaque filler lower than the second (col 3, ln 62-67), and a transition tube segment (172) that is interposed between the two other segments (col 4, ln 44-49) with the transition joint being variable along its length (col 4, 25-30) (Figures 3-10). Chee discloses the catheter being formed of several different materials (col 4, ln 3-17); in which physical parameters of the catheter can be varied along its length can be varied along its length along with the amount of radiopaque filler. The different disclosed tubing segments contain one or more lumens (Figures 3 and 10) that are integrally formed through solvent or heat joining (col 4, ln 20-30).

Application/Control Number: 10/728,267

Art Unit: 3763

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

- 1. Determining the scope and contents of the prior art.
- 2. Ascertaining the differences between the prior art and the claims at issue.
- 3. Resolving the level of ordinary skill in the pertinent art.
- 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 23-25 are rejected under 35 U.S.C 103(a) as being unpatentable over Chee et al. in view of Ju et al. (5,725,513). Chee et al. meets the claim limitations as described above except for a hub on the end of the catheter.

However, Ju et al. teaches a thin walled catheter with reinforcing sleeve.

Regarding claims 23-25, Ju et al. teaches a thin walled reinforced catheter with a hub that is configured to remain outside the body (20) (Figure 1).

At the time of the invention, it would have been obvious to add the catheter hub of Ju et al. to the catheter of Chee et al. because the addition of a catheter hub allows for agents to be delivery effectively through the catheter. The references are analogous in the art and with the instant invention; therefore, a combination is proper. Therefore,

Application/Control Number: 10/728,267

Art Unit: 3763

one skilled in the art would have combined the teachings in the references in light of the disclosure of Ju et al.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Koharski whose telephone number is 571-272-7230. The examiner can normally be reached on 7:30am to 4:00pm EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Date: /3//2/06

Christopher D. Koharski

Page 4

AU 3763